



Moldova

REQUEST FOR QUOTATION (RFQ)

NAME & ADDRESS OF FIRM:	DATE: 11 October 2011
	REFERENCE: RfQ11/00405

Dear Sir / Madam:

You are kindly requested to submit your quotation for the following items before **16:30 (local time) on 2 November 2011**.

Item	Generic Description	Quantity
LOT 1		
1.	<p>Vacuum tissue processor <u>Type:</u> vacuum, fully enclosed <u>Principle of functioning:</u> filling and draining by vacuum / pressure functions, pneumatic / hydraulic circuits completely sealed <u>Number of containers:</u> min 10 reagents, 4 +1 paraffin <u>Volume:</u> 5±2.5 liters <u>Capacity :</u> 300 samples +/- 50 <u>Temperature range:</u> shall be adjustable 50-70°C for paraffin and shall be adjustable 35-45°C for reagents, precision shall be ±1°C <u>Number protocols:</u> at least 12 <u>User security:</u> fully enclosed system, at least 2 carbon filters <u>Sample security:</u> various alarm systems, remote alarm monitoring interface <u>Vacuum</u> from 600mbars to 1200mbars <u>Cleaning functions</u> <u>Delay start function</u> At least 15 inch color <u>touch screen display</u> <u>Connections:</u> USB and printer Possibility to save data via <u>USB port</u> <u>Fast processing possibility:</u> 1mm biopsies in no more than 50 min At least 4 <u>paraffin baths</u> <u>Graphic visualization</u> of the at least 30 last executed processes 230V/50/60Hz <u>Operating conditions:</u> +15°C- 40°C, RH 10%-80%</p>	1 (one) unit
LOT 2		
1.	<p>Paraffin embedding system Modular system: paraffin dispenser, cooling plate and thermoblock <u>Paraffin dispenser</u> - At least 5 liters reservoir capacity with electronic temperature control shall be adjustable from 30 to 80° C, precision shall be ±1°C - Working surface at least 480±30 X 330±30 mm - At least 80 cassettes processing capacity - Waste collecting bottle and LED illumination - Maximum dimensions: 490±30 X 630±30 X 330±30 <u>Cooling plate</u> - Made of stainless steel - Cool down to 15° C at least 80 cassettes - Working surface at least: 320±20 X 420±20mm <u>Thermoblock</u> - Heated chamber with temperature control from 30 to 80° C and a bath for at least 150 cassettes with electronic temperature control form 30 to 80° C - 230V/50 Hz - Net weight no more than 70±10 kg</p>	1 (one) unit

LOT 3

<p>1.</p>	<p>Chemical Hood Type I</p> <p>The chemical extraction hood should guarantee operator's safety against chemical risk and the risk of fire or explosions.</p> <p><u>Construction characteristics:</u> Double rear wall forming the extraction chamber; Aerodynamic hood compartment to avoid vortex formation; 1 top air intake and 1 localized air intake on the worktop, guaranteeing fast evacuation of the hood compartment, safety system against explosion.</p> <p>The extractor unit must be equipped with a watertight motor IP 55 CH with polypropylene scroll and impeller, must have dimensions complying with the regulations, calculated according to the dimensions of the hood, the expulsion path, the different working processes and the products used.</p> <p><u>Standard equipment:</u> Glazed front door with latch and sliding glass panels; Glazed side walls; 1 electric control panel IP 667; 2 electric plugs 220V-16A; 1 cold water tap; 1 gas tap; 1 drainage basin; 1 lamp for internal lighting; 1 pull-out unit under the worktop, with hinged doors; 1 top expulsion fan; Internal water, gas, drainage and electric systems.</p> <p>The hood must be built according to technical standards UNI-FIN-DVGW-CI regarding laboratory furnishings.</p> <p><u>Dimensions:</u> mm 1480±130 x 8350±50 x 2180 ±130h; <u>Working surface:</u> chemical resistant polypropylene.</p>	<p align="center">3 (three) units</p>
<p>2.</p>	<p>Chemical Hood Type II</p> <p>The chemical extraction hood should guarantee operator's safety against chemical risk and the risk of fire or explosions.</p> <p><u>Construction characteristics:</u> Double rear wall forming the extraction chamber; Aerodynamic hood compartment to avoid vortex formation; 1 top air intake and 1 localized air intake on the worktop, guaranteeing fast evacuation of the hood compartment, safety system against explosion.</p> <p>The extractor unit must be equipped with a watertight motor IP 55 CH with polypropylene scroll and impeller, must have dimensions complying with the regulations, calculated according to the dimensions of the hood, the expulsion path, different working processes and the products used.</p> <p><u>Standard equipment:</u> Glazed front door with latch and sliding glass panels; Glazed side walls; 1 electric control panel IP 667; 2 electric plugs 220V-16A; 1 cold water tap; 1 gas tap; 1 drainage basin; 1 lamp for internal lighting; 1 pull-out unit under the worktop, with hinged doors; 1 top expulsion fan; Internal water, gas, drainage and electric systems.</p> <p>The hood must be built according to technical standards UNI-FIN-DVGW-CI regarding laboratory furnishings.</p> <p><u>Dimensions:</u> mm 2050±100 x 850±50 x 2480 ±130h; <u>Working surface:</u> chemical resistant polypropylene.</p>	<p align="center">1 (one) unit</p>



CONDITIONS	
Delivery Term (INCOTERMS 2000) & Place	<input type="checkbox"/> FCA <input type="checkbox"/> CPT <input checked="" type="checkbox"/> DAP
Delivery Place	8, Korolenko Street, MD-2025 Chisinau, Republic of Moldova
Payment Terms	100% upon delivery
Validity of Quotation	<input type="checkbox"/> 30 DAYS <input checked="" type="checkbox"/> 60 DAYS
Preliminary Examination - Completeness of quotation.	<input checked="" type="checkbox"/> Partial bids permitted (by LOTS) <input type="checkbox"/> Partial bids not permitted
Quantity change	The UNDP reserves the right to modify the quantity by 25% of the tendered goods
General Terms and Conditions	UNDP General Terms and Conditions for Goods/Services http://www.undp.org/procurement/operate.shtml

Please state	
Quantity discount and early payment discount	
Name and address of authorized service in Moldova	
Details on any warranty/guarantee conditions	
Details on delivery period	

REQUIREMENTS	
<u>Language:</u> All documentation, including installation and operating manuals shall be in:	
<input checked="" type="checkbox"/> English	<input type="checkbox"/> French <input type="checkbox"/> Spanish <input checked="" type="checkbox"/> Others: Romanian or Russian

QUOTATIONS/OFFERS PRESENTATION REQUIREMENTS:

The quotation/offer shall contain the following:

- Company profile (info up to 3 pages);
- Detailed description of the offered goods, including photos;
- Quotation for each item in MDL, EUR or USD exclusive of VAT, excise and other duties or taxes. (Evaluation shall be done in USD. Other currencies shall be converted into USD at the UN Operational Rate of Exchange on the day of competition deadline. Contract shall be signed with the winner in the currency of the offer. Payment to local suppliers will be done in MDL at the UN Operational Rate of exchange on the day of payment.);
- Additional information as requested under the "Please state" section;
- Statement on adherence to UNDP General Terms & Conditions and Payment & Delivery Terms above;
- Certificates of quality for the offered equipment;
- Statement or certificate of origin for the offered equipment;
- Offers shall be presented in Romanian, Russian or English.

MINIMUM QUALIFICATION REQUIREMENTS:

- Minimum 3 years of experience in the field;
- Availability of certificates of quality and origin for the offered equipment;
- Adherence to UNDP General Terms & Conditions and Payment & Delivery Terms above;
- Availability of authorised service in Moldova;
- The supplier shall install the equipment and provide induction trainings to the staff at no additional cost;
- Minimum warranty period of 24 months from the date of commissioning of the equipment;
- Availability to supply spare parts for at least 7 years from the date of commissioning of the equipment;
- User manuals for the equipment shall be presented in English with a certified translation into Romanian;
- Maximum delivery period not to exceed 45 calendar days upon signature of contract.

Offers will be evaluated based on their responsiveness to the technical specifications and the minimum qualification requirements, within the "either in or out" rule.



NAME, FUNCTIONAL TITLE: **Matilda Dimovska, Deputy Resident Representative**

Signature:  DATE: 13.10.2011

CONTACT PERSON: **Alexandru Cocirță, Project Manager** (alexandru.cocirta@undp.org)

CONTACT ADDRESS: UNDP Moldova, 131, 31 August 1989 Street, MD-2012 Chisinau

SUBMISSION OF OFFERS:

Offers shall be marked with the note "**RfQ11/00405: Forensic Laboratory Equipment**".

Offers shall reach the UNDP office not later than **2 November 2011, 16:30 (local time)**.

Offers can be submitted either in hard copy, or electronically. Offers received by fax will be rejected. Incomplete offers shall not be examined.

- a) Documents/offers in hard copy need to be submitted in a sealed envelope and addressed to:
**UNDP Moldova,
131, 31 August 1989 Street, MD-2012 Chisinau
Attention: Registry Office/Procurement**
- b) Offers sent electronically need to be addressed to the following e-mail address:
tenders-Moldova@undp.org