

RfQ11/00405: Forensic Laboratory Equipment

Corrections to Technical Specifications, Questions and Answers (as of 24 October 2011)

Corrections to Technical Specifications:

Take note of the following correction to technical specifications for **LOT 3. Item 1. Chemical Hood Type I:**

The text "Dimensions: mm 1480±130 x **8350**±50 x 2180 ±130h" is hereby corrected to now read as follows: "Dimensions: mm 1480±130 x **850**±50 x 2180 ±130h"

Questions and Answers:

Q1: The following parameters for hoods are in technical requirements:

1. "Hood Type 1 Dimensions: mm 1480 +/-130 x 8350 +/-50 x 2180 +/-130h"
2. "Hood Type 1 Dimensions: mm 2050 +/-100 x 850 +/-50 x 2480 +/-130h"

and we would like to clarify which of this sizes is depth, width and height?

A1: The dimensions are as follows:

1. "Hood Type 1 Dimensions: mm 1480 +/-130 (Length) x 850 +/-50 (Width) x 2180 +/-130 (Height)"
2. "Hood Type 1 Dimensions: mm 2050 +/-100 (Length) x 850 +/-50 (Width) x 2480 +/-130 (Height)"