

Due to a mechanical error, there are some minor changes to the technical specifications of the following item: Lot 2/Item 9 – Rotary evaporator.

The revised technical specifications shall be as follows:

9	<b>Rotary evaporator</b>		1
	Application	Sample preparation	
	Rotation speed	20-270rpm	
	Speed set	Digital set of rotation speed with display	
	Lift	Motor lift 18mm/sec	
	Heating bath temperature range	20-180°C	
	Over heat protection	Yes	
	Heating bath volume	4-5 liters	
	Digital display of vapor temperature	Yes	
	Timer	Automatic cut off at the end of distillation	
	Remote control	Yes	
	Protection class	IP 20	
	<b>Supplied glassware</b>	<b>Vertical condenser</b>	
Modular design with waterbath, vacuum pump and operating panel	Yes		